

Table 11. Post-Task Time Willing to Spend Estimates

<b>Measuring Concept</b>	<b>Subj. 3</b>	<b>Subj. 4</b>	<b>Subj. 6</b>	<b>Subj. 7</b>	<b>Subj. 8</b>	<b>Subj. 9</b>	<b>Subj. 11</b>	<b>Subj. Avg.</b>
Battery Attachment Task	1 m 30 s	NC	1 m	1 m 18 s	*	5 m	1 m	1 m 58 s
Battery Detachment Task	1 m	NC	1 m	15 s	*	5 m	1 m	1 m 39 s
Column Attachment Task	2 m	NC	5 m	2 m	*	5 m	5 m	3 m 48 s
Column Detachment Task	1 m	NC	5 m	2 m	*	3 m	5 m	3 m 12 s
Transfer Into Wheelchair Task	30 s	30 s	5 m	1 m	Not Tested	5 m	10 s	2 m 2 s
Transfer Out of Wheelchair Task	30 s	45 s	5 m	1 m	Not Tested	3 m	10 s	1 m 44 s
Switching Column Transfer to Manual	NC	30 s	5 m	20 s	Not Tested	3 m	5 s	1 m 47 s
Switching Column Manual to Power	20 s	5 s	5 m	15 s	Not Tested	3 m	5 s	1 m 28 s
Switching Column Power to Transfer	30 s	NC	5 m	15 s	Not Tested	3 m	5 s	1 m 46 s
Driving Clockwise	2 m	30 s	5 m	30 s	Not Tested	2 m	30 s	1 m 45 s
Driving Counterclockwise	2 m	30 s	5 m	30 s	Not Tested	2 m	30 s	1 m 45 s
Maneuvering Forward Task	1 m 30 s	40 s	5 m	17 s	Not Tested	1 m 30 s	30 s	1 m 35 s
Maneuvering in Reverse Task	5 m	45 s	NC	28 s	Not Tested	NC	1 m	1 m 48 s

Note: NC denotes not able to complete. m = minutes; s = seconds

\* Refused to give a specific time.